



## CERTIFICATE

Of Conformity
EU Council Directive 2014/30/EU
Electromagnetic Compatibility

Registration No.: ATSZAHE190325004

Report No.: SZAHE190325004-01

Applicant : HUIZHOU JINGHAO MEDICAL TECHNOLOGY CO., LTD

FLOOR 6, HUICHENG INDUSTRY BUILDING, NO.9 HUIFENG DONG'ER ROAD, ZHONGKAI HIGH-TECH ZONE,

HUIZHOU, GUANGDONG, CHINA

Product : hearing aid

Identification : Model No. : JH-351, JH-D26, JH-338, HA-70, HA-75,

JH-A39, JH-A17, JH-339

Trademark : N.A.

Rating : The adapter: DC5V, 400mA

hearing aid: DC3.7V, 40mAh

Test Standards : EN 55014-1: 2017

EN 55014-2: 2015

The certificate of conformity is based on an evaluation of a sample of the above-mentioned product. Technical report and documentation are at the applicant's disposal. This is to certify that the tested sample is in conformity with all provisions of Annex II of Council Directive 2014/30/EU, in its latest amended version, referred to EMC Directive. The certificate does not imply assessment of the production and does not permit the use of Lab's logo. The applicant of the certificate is authorized to use this certificate in connection with EU declaration of conformity to Article 15 of the Directive.

Anbotek
Product Safety

\* Approved \*

Certified by

sally zhane

Sally Zhang Manager

Apr. 03, 2019 Date

CE

The CE Marking may only be used if all relevant and effective EU Directives are complied with

Shenzhen Anbotek Compliance Laboratory Limited

1/F, Building D, Sogood Science and Technology Park, Sanwei community,

Hangcheng Street, Bao'an District, Shenzhen, Guangdong, China.518102
Tel: (86)755-26066440
Http://www.anbotek.com
Email: service@anbotek.com